

Advisory Notice

Date Released: 2018-January-10

Posted By: Quality Control

Product Affected: Vitamin D Direct

Brief Description of Advisory: Control Now Included, See Update IFU

To whom it may concern,

Based on market feedback and customer request, Monobind is now including a two-level control in its Vitamin D Direct kits, per the following:

- Vit D Direct AccuBind® ELISA Item #: 9425-300
- Vit D Direct AccuLite® CLIA Item #: 9475-300

The product Instructions for Use (**IFU**) have been updated to a new version to reflect this change and can be downloaded at <http://www.monobind.com/Support-Areas/product-inserts>

The **Design Output** has also been updated and is available by request.

Product Price: There has been a slight increase to offset these new reagents and current customers will be notified.

As an official of the company, I verify this information to be true and correct.

Sincerely,

Anthony Shatola



Quality Control Director