

# Advisory Notice

**Date Released: 2018-December-4**

**Posted By: Quality Control**

**Product Affected: Free Testosterone**

**Brief Description of Advisory: Larger Measurement Range, Control Now Included, Antibody Plate, See Updated IFU**

---

To whom it may concern,

Based on market feedback and customer request, Monobind has revised its Free Testosterone assays to have a larger measurement range (0-60 pg/ml from 0-37.5 pg/ml), a three-level control, as well as an antibody-coated plate which affects the following products:

- Free Testosterone AccuBind® ELISA                      Item #: 5325-300
- Free Testosterone AccuLite® CLIA                      Item #: 5375-300

This applies all lots produced in December 2018 (going forward), specifically LOT 53K1L8 and later.

The product Instructions for Use (IFU) have been updated to a new version to reflect this change and can be downloaded at <http://www.monobind.com/Support-Areas/product-inserts>

The **Design Output** has also been updated and is available by request.

**Product Price:** There has been no price increase.

As an official of the company, I verify this information to be true and correct.

Sincerely,

Anthony Shatola



Quality Control Director